



Borehole **22-00-02**

Log Event A

Borehole Information

Farm : <u>BY</u>	Tank : <u>BY</u>	Site Number : <u>299-E33-240</u>
N-Coord : <u>46,072</u>	W-Coord : <u>53,195</u>	TOC Elevation : <u>648.05</u>
Water Level, ft :	Date Drilled : <u>12/7/1973</u>	

Casing Record

Type : <u>Steel-welded</u>	Thickness : <u>0.280</u>	ID, in. : <u>6</u>
Top Depth, ft. : <u>0</u>	Bottom Depth, ft. : <u>100</u>	

Borehole Notes:

The borehole was drilled with a cable tool drilling rig, and the casing is apparently ungrouted and unperforated.

Equipment Information

Logging System : <u>1</u>	Detector Type : <u>HPGe</u>	Detector Efficiency: <u>35.0 %</u>
Calibration Date : <u>03/1995</u>	Calibration Reference : <u>GJPO-HAN-1</u>	Logging Procedure : <u>P-GJPO-1783</u>

Log Run Information

Log Run Number : <u>1</u>	Log Run Date : <u>8/4/1995</u>	Logging Engineer: <u>Bob Spatz</u>
Start Depth, ft.: <u>99.0</u>	Counting Time, sec.: <u>100</u>	L/R : <u>L</u> Shield : <u>N</u>
Finish Depth, ft. : <u>36.0</u>	MSA Interval, ft. : <u>0.5</u>	Log Speed, ft/min.: <u>n/a</u>
Log Run Number : <u>2</u>	Log Run Date : <u>8/7/1995</u>	Logging Engineer: <u>Alan Pearson</u>
Start Depth, ft.: <u>0.0</u>	Counting Time, sec.: <u>100</u>	L/R : <u>L</u> Shield : <u>N</u>
Finish Depth, ft. : <u>37.0</u>	MSA Interval, ft. : <u>0.5</u>	Log Speed, ft/min.: <u>n/a</u>



Borehole

22-00-02

Log Event A

Analysis Information

Analyst : D.C. Stromswold

Data Processing Reference : P-GJPO-1787

Analysis Date : 1/23/1996

Analysis Notes :

Verification spectra collected before and after the log runs showed that the logging tool was operating properly.

Gain drifts were minimal, enabling a single energy calibration to be used during data processing for each run.

Repeatability at the overlap logging section was within the calculated statistical uncertainties.

Correction factors for 0.25-in.-thick steel casing were used during data processing. No water correction was applied because the borehole was dry.

Cs-137, Co-60, and Sb-125 were the man-made contaminants detected in this borehole. Cs-137 was detected mainly above 20 ft. Co-60 was detected mainly in two intervals: one near 52 ft with concentrations near the minimum detection level, and the other near 85 ft with concentrations up to 2 pCi/g. Sb-125 detection was limited to two depths near 52 ft.

The K-40 concentration increases below a depth of about 47 ft, which coincides with the bottoms of the nearby tanks.

See the Tank Summary Data Report for BY-103 for additional log analysis.

Log Plot Notes:

Separate log plots show the man-made (e.g., Cs-137) and the naturally occurring radionuclides (K-40, U-238, and Th-232). The natural radionuclides can be used for lithology interpretations. The headings of the plots identify the specific gamma rays used to calculate the concentrations.

A combination plot includes both the man-made and natural radionuclides, in addition to the total gamma derived from the spectral data and the Westinghouse Hanford Company (WHC) Tank Farms gross gamma log. The gross gamma plot displays the latest available digital data from WHC with no attempt to adjust the depths to coincide with the SGLS data.

Uncertainty bars on the plots show the statistical uncertainties for the measurements as 95-percent confidence intervals. Open circles on the plots give the minimum detection level (MDL). The MDL of a radionuclide represents the lowest concentration at which positive identification of a gamma-ray peak is statistically defensible.